

330桃園市桃園區大業路一段18號6樓之1

6F.-1, No.18, Sec. 1, Daye Rd., Taoyuan Dist., Taoyuan City 330, Taiwan (R.O.C.)

電話Tel: 03-331-2677 e-mail: sales@ubiq.com.tw 傳真Fax: 03-335-7412 Website: www.ubiq.com.tw

Battery Automation Test Systems

Product name: Battery Automation test systems Model: BAT-750B 5V (5mA、50mA two-stage) 8 Channels

Specifications:



- 1. Charge: CC \ CV (0.005V~4.7V)
- 2. Discharge: CC、CP、CV (0.005V Constant voltage discharge)
- 3. Current accuracy: $5mA \pm (0.06\% + 1uA) & 50mA \pm (0.06\% + 10uA)$
- 4. Voltage accuracy: $\pm (0.05\% + 2mV)$
- 5. Including AC internal resistance measurement specifications:
- $0\sim400$ m $\Omega\pm0.2$ m Ω
- 6. The computer is included and the latest version testing software is installed
 - and calibrated for immediate use.
- 7. Including 8 sets of CR2032 test fixtures and 8 sets of four-wire charging and discharging wires
- 8. Data reading speed: 60ms