



優必克科技股份有限公司
UBIQ Technology Co., Ltd.

330桃園市桃園區大業路一段18號6樓之1

6F.-1, No.18, Sec. 1, Daye Rd., Taoyuan Dist., Taoyuan City 330, Taiwan (R.O.C.)

電話Tel : 03-331-2677 e-mail: sales@ubiq.com.tw

傳真Fax : 03-335-7412 Website: www.ubiq.com.tw

充放電機 (8 channels)

型號：BAT-750B



規格：

1. 充電：CC、CV (0.005V~4.7V)
2. 放電：CC、CP、CV□ (0.005V 定電壓放電)
3. 電流精準度：5mA±(0.06%+1uA) & 50mA±(0.06%+10uA)
4. 電壓精準度：±(0.05%+2mV)
5. 含交流內阻量測規格：0~400mΩ±0.2mΩ□
6. 含控制電腦一台及測試軟體一套
7. 含8組CR2032測試治具及8組四線式充放電線
8. 資料讀取速度60ms



優必克科技股份有限公司
UBIQ Technology Co., Ltd.

330桃園市桃園區大業路一段18號6樓之1

6F.-1, No.18, Sec. 1, Daye Rd., Taoyuan Dist., Taoyuan City 330, Taiwan (R.O.C.)

電話Tel : 03-331-2677 e-mail: sales@ubiq.com.tw

傳真Fax : 03-335-7412 Website: www.ubiq.com.tw

Battery Automation Test Systems

Product name : Battery Automation test systems
Model : BAT-750B 5V (5mA、50mA two-stage) 8 Channels



Specifications:

1. Charge: CC 、 CV (0.005V~4.7V)
2. Discharge: CC、 CP、 CV (0.005V Constant voltage discharge)
3. Current accuracy: $5\text{mA} \pm (0.06\% + 1\mu\text{A})$ & $50\text{mA} \pm (0.06\% + 10\mu\text{A})$
4. Voltage accuracy: $\pm (0.05\% + 2\text{mV})$
5. Including AC internal resistance measurement specifications :
 $0 \sim 400\text{m}\Omega \pm 0.2\text{m}\Omega$
6. The computer is included and the latest version testing software is installed
and calibrated for immediate use.
7. Including 8 sets of CR2032 test fixtures and 8 sets of four-wire charging and discharging wires
8. Data reading speed: 60ms